



## Declaration of Conformity

For the following equipment :

Product Name: AC/DC Switching Adapter

Model Designation: NGE12xyzzzz, NGE18xyzzzz (x=l, UK, y=05, 09, 12, 15, 18, 24, zzzz=maybe Blank, -, 0-9, A-Z or a-z for market purpose)

The designated product(s) is(are) in conformity with the relevant legislation:

**The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012:** SI 2012 No. 3032

**Electrical Equipment (Safety) Regulations 2016 :**

BS EN 62368-1:2014+A11:2017

Dekra Certificate: 35-133375

BS EN 60335-1:2012+A15:2021

Dekra Certificate: 35-134255

BS EN IEC 61558-1:2019 ;BS EN 61558-2-16:2009/A1:2013

Dekra Certificate: 35-134253

**Medical Devices Regulations 2002 (SI 2002 No 618) (UK MDR 2002)**

BS EN 60601-1:2006+A2:2021 ; BS EN 60601-1-11:2015+A1:2021

Dekra Certificate: 35-134254

BS EN 60601-1-2:2015+A1:2021

**Electrical Compatibility Regulations 2016 :**

**EMI (Electro-Magnetic Interference)**

Conducted emission BS EN 55032:2015+A1:2020

Radiated emission BS EN 55032:2015+A11:2020

BS EN 55011:2016+A2:2021 Class B

Harmonic current BS EN IEC 61000-3-2:2019+A1:2021 Class A

Voltage flicker BS EN 61000-3-3:2013+A1:2019 Clause 5

**EMS (Electro-Magnetic Susceptibility)**

BS EN 55035:2017+A11:2020 BS EN IEC 61204-3:2018 BS EN 60601-1-2:2015+A1:2021

ESD air BS EN 61000-4-2:2009 Level 4 15KV

RF field susceptibility BS EN IEC 61000-4-3:2020 Level 2 3V/m(80MHz~2.7GHz)

RF field susceptibility BS EN IEC 61000-4-3:2020 Table 9 9~28V/m (385MHz~5.78GHz)

EFT bursts BS EN 61000-4-4:2012 Level 3 2KV

Surge susceptibility BS EN 61000-4-5:2014+A1:2017 Level 3 1KV/Line-Line

Conducted susceptibility BS EN 61000-4-6:2014 Level 2 3V

Magnetic field immunity BS EN 61000-4-8:2010 Level 4 30A/m

Voltage dip, interruption BS EN IEC 61000-4-11:2020 0% residual voltage for 0.5 cycles , 70% residual voltage for 25 cycles , 0% residual voltage for 250 cycles

**Note:**

The power supply is considered as a component that will be operated in combination with final equipment. Since EMC performance will be affected by the complete installation, the final equipment manufacturers must re-qualify EMC Directive on the complete installation again.

For guidance on how to perform these EMC tests, please refer to TDF (Technical Documentation File).

The product under declaration is just a unit without medical function. Complete MDR should only be verified when it is used together with particular medical device(s).

**The Ecodesign for Energy-Related Products and Energy Information (Amendment) (EU Exit) Regulations 2020**

This Declaration is effective from serial number SC3xxxxxxx

Person responsible for marking this declaration :

MEAN WELL Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 24891, Taiwan

(Manufacturer Address)

Aries Jian/ Director, Group R&D :

(Name / Position)

*Aries*  
(Signature)

Alex Tsai/Director, Product Strategy Center :

(Name / Position)

*[Signature]*  
(Signature)

Taiwan

(Place)

Dec. 29th, 2023

(Date)