

# Technical Compliance Statement

## EMC Test Report

For the following information


Ref. File No.: C1M2308278

Product : Medical Type Switching Power Supply  
Model Number : RPS-300-x-y (x=12, 15, 24, 27, 48; y=C or Blank)  
Applicant : MEAN WELL Enterprises Co., Ltd.  
Manufacturer #1 : MEAN WELL Enterprises Co., Ltd.  
Manufacturer #2 : MEAN WELL (GUANGZHOU) ELECTRONICS CO.,  
LTD HUADU BRANCH.  
Manufacturer #3 : SuZhou MEAN WELL Technology Co., Ltd.  
Manufacturer #4 : YONGDEN YECHNOLOGY CORPORATION  
Manufacturer #5 : MEAN WELL INDIA ELECTRONICS PRIVATE  
LIMITED  
Standards :  
EN 60601-1-2:2015 +A1:2021  
Emission: EN 55011:2016 +A11:2020  
EN IEC 61000-3-2:2019 and EN 61000-3-3:2013 +A1:2019  
Immunity: EN 61000-4-2:2009, EN IEC 61000-4-3:2020,  
EN 61000-4-4:2012, EN 61000-4-5:2014 +A1:2017, EN 61000-4-6:2014,  
EN 61000-4-8:2010, EN IEC 61000-4-11:2020

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the Regulation (EU) 2017/745 concerning medical device.

The test data and results are issued on the test report no. **EM-E160247C**.

Signature



Alex Deng/Deputy Manager  
Date: 2023. 09. 13

Test Laboratory:  
Audix Technology Corporation, EMC Department  
TAF Accreditation No.: 1724  
Web Site: [www.audixtech.com](http://www.audixtech.com)

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.